

EU Type Examination Certificate

This is to certify that:

WRP Asia Pacific SDN BHD
Lot 1, Jalan 3
Kawasan Perusahaan
Bandar Baru Salak Tinggi
Sepang, Selangor Darul Ehsan
Selangor
43900
Malaysia

Holds Certificate Number:

CE 688314

In respect of:

Powder Free Nitrile Examination Gloves (WRP NBR 3549)
To EN 388:2016, EN 374-1:2016, EN 374-2:2014, EN 374-5:2016, EN 16523-1:2015, and EN 420:2003+A1: 2009
See continuation sheet for details

on the basis BSI carried out the relevant Type Examination procedures in accordance with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V

For and on behalf of BSI:


Chris Lewis - Certification Director, Product Certification

First Issued: 2018-04-26

Latest Issue: 2018-04-26

Effective Date: 2018-04-26

Expiry Date: 2023-04-26

Page: 1 of 4

...making excellence a habit.™

EU Type Examination Certificate

No. CE 688314

Product Specification

Model: Nitrile disposable gloves (WRP NBR 3549)

Classification: Protective glove for use against chemical and micro-organism hazards.

Description: Powder Free Nitrile Disposable Protective Gloves manufactured from 100% nitrile synthetic rubber latex (Acrylonitrile-butadiene) not containing natural rubber latex, to eliminate Type I allergic reactions. Inner surface of gloves is smooth surface that assists in donning the gloves without using lubricant such as powder on the glove surface. The glove is available as either Sterile or Non-Sterile as indicated.

The main features of this protective glove are:

- Beaded cuff
- Micro-organisms penetration resistance
- Chemical permeation resistance

Models in Range: Nitrile disposable gloves.

PPE Category: Category III – complex

Applicable standards:

EN 420:2003+A1:2009 - General requirements for gloves

EN 388:2016 - Mechanical Hazards

EN 374-1:2016 - Protective gloves against chemicals and micro-organisms

EN 374-2:2014 - Resistance to penetration

EN 16523-1:2015 - Additional chemicals tested

EN 374-5:2016 - Terminology and performance requirements for micro-organisms

First Issued: 2018-04-26

Latest Issue: 2018-04-26

Effective Date: 2018-04-26

Expiry Date: 2023-04-26

Page: 2 of 4

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

EU Type Examination Certificate

No. CE 688314

Model: **Nitrile disposable gloves (WRP NBR 3549)**

Available in sizes XS, S, M, L and XL

Performance: **General requirements for gloves to EN 420:2003+A1:2009**

Dexterity: 5

Mechanical Hazards to EN 388:2016

Abrasion resistance – level 0

Blade cut resistance – level 0

Tear resistance – level 0

Puncture resistance – level 0

Resistance to penetration to EN 374-2:2014

Pass

**Resistance to chemical permeation to EN 374-1:2016
(test method EN 16523-1:2015)**

n-Heptane (J) – level 1

40% Sodium hydroxide (K) – level 6

96% Sulphuric acid (L) – level 0

37% Formaldehyde (T) – level 5

30 % Hydrogen peroxide (P) – level 5

**Resistance to penetration by blood-borne pathogens
tested to EN 374-5:2016 (test method EN 16604:2014)**

Protection against bacteria and fungi
Protection against viruses

Pass
Pass

First Issued: 2018-04-26

Latest Issue: 2018-04-26

Effective Date: 2018-04-26

Expiry Date: 2023-04-26

Page: 3 of 4

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 688314

Certificate Administration Details

Technical File Reference: TF/PPE/002

Certificate Amendment Record

Issue date	Comments	BSI Internal report No.
April 2018	First issue.	0086:18:8888423

Note: The Certificate holder is responsible for keeping the Notified Body advised of changes to any aspect of the overall process used in the manufacture of the product.

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The Conformity to Type Based on Quality Assurance of the Production Process, Annex VIII (Module D), for the product are referenced in BSI issued Certificate number CE 51699

First Issued: 2018-04-26

Latest Issue: 2018-04-26

Effective Date: 2018-04-26

Expiry Date: 2023-04-26

Page: 4 of 4

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.